

LigaTrap® Rat IgM Loose Resin Protocol

Introduction

LigaTrap® Rat IgM Prepacked Columns are capable of binding ≥ 10 mg monoclonal Rat IgM /ml Resin. Kappa and Lambda IgM can be purified using this product. LigaTrap® IgM purification resin is capable of processing and purifying monoclonal antibodies from cell culture supernatant, ascites fluid, hybridoma, and other sources of recombinant IgM. **Serum applications are not recommended with all LigaTrap® IgM Purification products, due to potential cross reactivity with other immunoglobulins.**

Chromatographic Buffers:

Part #	Name	Composition
BU-131-FP	LigaTrap Sample Diluent 2.0	50mg/ml Adipic Acid, 4.0M NaCl, pH 5.8
BU-132-FP	LigaTrap Equilibration/Wash Buffer 2.0	10mg/ml Adipic Acid, 800mM NaCl, pH 5.8
BU-133-FP	LigaTrap IgM Elution Buffer	0.05M Sodium Acetate, pH 3.8
BU-124-FP	LigaTrap Regeneration Buffer	0.1M Glycine, pH 2.5
BU-125-FP	LigaTrap Neutralization Buffer	3.0M Tris-Base, pH 11.1
BU-126-FP	LigaTrap Storage Buffer	10mM Sodium Phosphate, 0.15M NaCl, 0.05% Sodium Azide, pH 7.2

Note: Adipic Acid can be purchased from Sigma (A26357-500G)

Note: Adipic Acid is insoluble at low pH. It will solubilize as the pH increases to > 5.0 . For example, after adding all components of LigaTrap Sample Diluent 2.0 Buffer, a 500ml batch will require ~60mL of 5N NaOH to begin dissolving the adipic acid. Continue to titrate with NaOH until a final pH of 5.8 is achieved. Then QS to the desired volume.

Note: For best results, titrate LigaTrap® Elution Buffer with Glacial Acetic Acid

Note: To limit precipitation of Tris-Base, store LigaTrap Neutralization Buffer at room temperature.

Prepare Sample For Binding:

1. Add the **LigaTrap Sample Diluent 2.0** to the sample containing IgM at a ratio of 1:4
(For example: Add 2ml LigaTrap Sample Diluent 2.0 to 8ml of sample, or 200ml to 800ml of sample, etc...)

Chromatographic Protocol:

- Pack the appropriately sized column volume (CV), for your particular application. **Ensure not to exceed a maximum pressure of 0.2 MPa (2 Bar).**
 - Recommend 5-minute residence times for the Sample Load and all other chromatographic steps
 - *If using a FPLC system capable of setting multiple pressure alarms, set the pre-column pressure alarm to 0.2MPa.*
1. Equilibrate the resin with 10CV of **LigaTrap Equilibration/Wash Buffer 2.0 (BU-132-FP)**
 2. Load the "prepared" sample (described above)
 3. Wash the resin with 10-15CV of **LigaTrap Equilibration/Wash Buffer 2.0 (BU-132-FP)**
 4. Elute bound antibody with 10 CV of **LigaTrap IgM Elution Buffer (BU-133-FP)**. Note, for higher concentration antibody elute with 5 CV, but if higher yields are desired, use 10 CV. Neutralize with 12% v/v of **LigaTrap Neutralization Buffer**.
 5. Regenerate the column with 10CV of **LigaTrap Regeneration Buffer (BU-124-FP)**

6. Re-Equilibrate the resin with 10CV **LigaTrap Equilibration/Wash Buffer 2.0 (BU-132-FP)** if an additional purification cycle will be done.
7. If purification is complete, store the resin by running 5CV of **LigaTrap Storage Buffer (BU-126-FP)** over the column

Other LigaTrap® Products

Target Species	Isotype	Part Number		
		Prepacked Columns	5mL Loose Resin	Purification Spin Column Kit
Human	IgG	LT-095-1x1ml LT-095-3x1ml LT-095-1x5ml	LT-095	LT-095KIT
	IgM	LT-143-1x1ml LT-143-3x1ml LT-143-1x5ml	LT-143	LT-143KIT
	IgA	LT-146-1x1ml LT-146-3x1ml LT-146-1x5ml	LT-146	LT-146KIT
Mouse	IgG	LT-137-1x1ml LT-137-3x1ml LT-137-1x5ml	LT-137	LT-137KIT
	IgM	LT-145-1x1ml LT-145-3x1ml LT-145-1x5ml	LT-145	LT-145KIT
Rat	IgG	LT-138-1x1ml LT-138-3x1ml LT-138-1x5ml	LT-138	LT-138KIT
	IgM	LT-147-1x1ml LT-147-3x1ml LT-147-1x5ml	LT-147	LT-147KIT
Sheep	IgG	LT-141-1x1ml LT-141-3x1ml LT-141-1x5ml	LT-141	LT-141KIT
Llama	IgG	LT-144-1x1ml LT-144-3x1ml LT-144-1x5ml	LT-144	LT-144KIT
Goat	IgG	LT-136-1x1ml LT-136-3x1ml LT-136-1x5ml	LT-136	LT-136KIT
Rabbit	IgG	LT-139-1x1ml LT-139-3x1ml LT-139-1x5ml	LT-139	LT-139KIT
Chicken	IgY	LT-142-1x1ml LT-142-3x1ml LT-142-1x5ml	LT-142	LT-142KIT

For further product information please visit our website at LigaTrap.com. For technical support and questions email us at techsupport@ligatrap.com