

LigaTrap® Rat IgM Microspin Columns

Product Instructions

Introduction

LigaTrap Rat IgM Purification Resin is engineered to purify high quality Rat IgM antibodies from polyclonal and monoclonal sources. LigaTrap Rat IgM Purification Resin is capable of binding ≥ 10 mg Rat IgM /mL resin. Each microspin column may be used, regenerated, and reused up to 10 times with minimal loss in binding capacity. Kappa and Lambda IgM may be purified using this product. **Serum applications are not recommended with all LigaTrap IgM Purification products, due to potential cross reactivity with other immunoglobulins.**

Table 1. Product Contents

Part #	Item	Quantity
LT-147-MSC	Microspin Columns- centrifuge columns supplied with caps and plug. Each microspin column contains 0.1 mL LigaTrap Rat IgM Purification Resin in PBS buffer with 0.05% sodium azide.	10

Additional Materials Required

Buffers

All buffers can be prepared as shown in **Table 2** below, or can be purchased as pre-qualified buffers from the LigaTrap Technologies website.

Table 2. Chromatographic Buffers

Part #	Name	Composition
BU-131-FP	LigaTrap Sample Diluent 2.0	50mg/mL Adipic Acid, 4.0M NaCl, pH 5.8
BU-132-FP	LigaTrap Equilibration/Wash Buffer 2.0	10mg/mL Adipic Acid, 800mM NaCl, pH 5.8
BU-133-FP	LigaTrap IgM Elution Buffer	500mM Sodium Acetate, pH 3.8
BU-124-FP	LigaTrap Regeneration Buffer	0.1M Glycine, pH 2.5
BU-125-FP	LigaTrap Neutralization Buffer	3.0M Tris-Base, pH 11.1
BU-126-FP	LigaTrap Storage Buffer	10mM Sodium Phosphate, 0.15M NaCl, 0.05% Sodium Azide, pH 7.2

Note: Adipic acid is insoluble at low pH. It will solubilize as the pH increases to > 5.0 .

Note: For best results, titrate LigaTrap Elution Buffer with Glacial Acetic Acid

Note: To limit precipitation of Tris-Base, store LigaTrap Neutralization Buffer at room temperature.

Equipment

- Microcentrifuge set between 1000–3000 x g
- Vortex/Mixer
- Centrifugation tubes or container for sample preparations

Antibody Purification Procedure

Sample Preparation

1. In a separate tube add 320µL of sample matrix (i.e. hybridoma supernatant or cell culture fluid) containing Rat IgM.
2. Add 80µL of **LigaTrap Sample Diluent 2.0 (BU-131-FP)** to the sample. Mix briefly by vortexing.

Purification

3. Snap off the bottom plug on the microspin column. **Save this plug, as it will be needed to stopper the column.**
4. Insert the microspin column into a 2.0 mL collection tube. Equilibrate resin by adding 400µL of **LigaTrap Equilibration/Wash Buffer 2.0 (BU-132-FP)** to the unplugged microspin column. Centrifuge between 1000-3000 x g for 1 minute. Discard the buffer in collection tube. Repeat for two additional 400µL equilibrations. Insert the bottom plug onto the microspin column.
5. Transfer 400µL of the prepared sample (from Step # 2) to the equilibrated column. Place screw cap on snugly. Continue to mix/shake the sample and resin continuously for **5 minutes**. Remove bottom plug and insert microspin column into a new 2.0 mL collection tube. Centrifuge between 1000-3000 x g for 1 minute. Retain flow through.
6. Insert bottom plug onto the microspin column and add 400µL of the **LigaTrap Equilibration/Wash Buffer 2.0 (BU-132-FP)**. Mix/shake resin continuously for 5 minutes. Remove bottom plug and insert microspin column into a new 2.0 mL collection tube. Centrifuge between 1000-3000 x g for 1 minute. Retain wash flow through. Repeat process for a second 400µL wash.
7. Insert the bottom plug onto the microspin column and add 400µL of **LigaTrap IgM Elution Buffer (BU-133-FP)**. Mix/shake resin continuously for 5 minutes. Remove bottom plug and insert the microspin column into a new 2.0 mL collection tube labeled **Eluate 1**. Centrifuge between 1000-3000 x g for 1 minute. Repeat process for a second 400µL elution and use a new 2.0 mL collection tube labeled **Eluate 2**.
Note: The eluates contain the purified antibodies. **Do not discard!**
8. Add 50µL (12.5% v/v of elution samples) of **LigaTrap Neutralization Buffer (BU-125-FP)** to each of the eluates obtained in Step # 7. Vortex briefly. The antibody will be near neutral pH and is ready for downstream applications.
Note: There are no preservatives in the antibody. Use the antibody within one week or aliquot and store at -20° C or colder. Avoid multiple freeze thaws.
9. Insert the bottom plug onto the microspin column and add 400µL of **LigaTrap Regeneration Buffer (BU-124-FP)**. Mix/shake resin continuously for 5 minutes. Remove bottom plug and insert microspin column into a new 2.0 mL collection tube. Centrifuge between 1000-3000 x g for 1 minute. Retain regeneration flow through.
10. Add 50µL of **LigaTrap Neutralization Buffer (BU-125-FP)** to the regeneration flow through obtained in Step # 9. Vortex briefly.

11. If the column will not be reused, it may be discarded. If the microspin column is to be reused, re-equilibrate the resin by repeating the process described in Step # 4.
12. To store resin, remove bottom plug and insert microspin column into a new 2.0 mL collection tube. Add 400µL of **LigaTrap Storage Buffer (BU-126-FP)**. Centrifuge between 1000-3000 x g for 1 minute. Repeat for two more 400µL washes. Once complete, insert the bottom plug onto the microspin column and add 400µL of fresh **LigaTrap Storage Buffer (BU-126-FP)**. Store plugged microspin column upright at 2-8° C.

Other LigaTrap Products:

Target Species	Antibody	Part Number			
		Loose Resin	Microspin Columns	Prepacked Columns	Purification Kits
Human	IgG	LT-095	LT-095-MSC	LT-095-1x1mL LT-095-3x1mL LT-095-1x5mL	LT-095KIT LT-095-1mL KIT LT-095-5mL KIT
	IgM	LT-143	LT-143-MSC	LT-143-1x1mL LT-143-3x1mL LT-143-1x5mL	LT-143KIT LT-143-1mL KIT LT-143-5mL KIT
	IgA	LT-146	LT-146-MSC	LT-146-1x1mL LT-146-3x1mL LT-146-1x5mL	LT-146KIT LT-146-1mL KIT LT-146-5mL KIT
Mouse	IgG	LT-137	LT-137-MSC	LT-137-1x1mL LT-137-3x1mL LT-137-1x5mL	LT-137KIT LT-137-1mL KIT LT-137-5mL KIT
	IgM	LT-145	LT-145-MSC	LT-145-1x1mL LT-145-3x1mL LT-145-1x5mL	LT-145KIT LT-145-1mL KIT LT-145-5mL KIT
Rat	IgG	LT-138	LT-138-MSC	LT-138-1x1mL LT-138-3x1mL LT-138-1x5mL	LT-138KIT LT-138-1mL KIT LT-138-5mL KIT
	IgM	LT-147	LT-147-MSC	LT-147-1x1mL LT-147-3x1mL LT-147-1x5mL	LT-147KIT LT-147-1mL KIT LT-147-5mL KIT
Llama	IgG	LT-144	LT-144-MSC	LT-144-1x1mL LT-144-3x1mL LT-144-1x5mL	LT-144KIT LT-144-1mL KIT LT-144-5mL KIT
Goat	IgG	LT-136	LT-136-MSC	LT-136-1x1mL LT-136-3x1mL LT-136-1x5mL	LT-136KIT LT-136-1mL KIT LT-136-5mL KIT
Rabbit	IgG	LT-139	LT-139-MSC	LT-139-1x1mL LT-139-3x1mL LT-139-1x5mL	LT-139KIT LT-139-1mL KIT LT-139-5mL KIT
Chicken	IgY	LT-142	LT-142-MSC	LT-142-1x1mL LT-142-3x1mL LT-142-1x5mL	LT-142KIT LT-142-1mL KIT LT-142-5mL KIT

For further product information, please visit our website at www.LigaTrap.com. For technical support and questions, email us at info@ligatrap.com